BioFlo® 510 Initial Load Software

NOTE: This update is only to be used when installing a NEW CPU board!

Date/File/Version	Description
6-August-2013 **UPDATE_M1361_Initial_A7.exe** Version A7	BioFlo 510 software for initial installation.

Instructions for loading software on a new computer board

On a new computer board it will be necessary to restore the operating system. This is done automatically; the only requirements will be to calibrate the touchscreen and then to update the software to latest version.

During this process the screen may be blank or may appear as if the system is not updating; **do not restart the unit! This process will take about 10 minutes to complete.**

1. Install a USB external memory device into the computer's USB drive connection.
2. Go to the Saved file on your hard drive and double click the file.
3. Pop up menu comes up - Select Run.
4. From the pop up menu - under "Unzip to folder" - click "Browse" and select USB drive (Removable Disk).
5. Select Unzip button and Okay.
6. After download is complete select "Close" to close the dialog box.
7. Connect the USB thumb drive with latest software to USB connector on unit.
8. Turn power on.
9. After several minutes the calibration screen will be displayed. Complete the calibration process.

10. Once calibration is completed a software loader screen is displayed.

Select the appropriate software that needs to be loaded into the compact flash and click on Load Update *(Note: if available updates do not display, click "Refresh List")*.

11. Once the unit reboots remove the thumb drive with the software update.
12. Acknowledge all screens to setup process.
13. Select "Setup" then the "Hardware Setup" tab and add the hardware required by the unit.
14. Select "Setup" then the "Controller Setup" tab and select Save changes.