

# Test Report

## – Eppendorf Forensic DNA Grade –

according to ISO 18385

**Blank example certificate**

• <b>Human DNA</b> Method: <i>real-time</i> PCR acc. to SOP 111141/V4 Limit: < 0.5 pg/ $\mu$ L	• <b>DNase</b> Method: DNA Digestion acc. to SOP 100570/V4 Limit: Not detectable (LOD: $1.0 \times 10^{-7}$ Kunitz units)
• <b>PCR inhibition</b> Method: <i>real-time</i> PCR acc. to SOP 100559/V4 Limit: Less than 10 targets amplifiable	• <b>RNase</b> Method: RNA Digestion acc. to SOP 100571/V4 Limit: Not detectable (LOD: $1.0 \times 10^{-9}$ Kunitz units)

Customer: Eppendorf SE, Barkhausenweg 1, D-22339 Hamburg

Date of sample receipt / examination:

Testing of the rinse solution of the above described lot showed conformity within the limits of detection.

.....  
Date of release

.....  
Registration number

.....  
Dr. rer. nat. Sybille Kuhnert (Head)

The test results refer exclusively to the items tested. The test report must not be copied partially without the approval of LADR GmbH.