



UK Health  
Security  
Agency

# Certificate of Containment Testing

## Containment Testing of Rotor FA-24x2 with updated seal in an Eppendorf Bench Top 5425/R centrifuge

Report No. 24/001

**Report Prepared For:** Eppendorf Zentrifugen GmbH

**Issue Date:** 17 May 2024

### Test Summary

Rotor FA-24x2 was containment tested with an updated seal at 15,060 rpm in an Eppendorf bench top 5425/R centrifuge, using Annex AA of IEC 61010-2-020:2016 (3<sup>rd</sup> Ed.). The sealed rotor was shown to contain a spill.

**Report Written By**

**Name:** Ms Helen Hookway

**Title:** Biosafety Scientist

**Report Authorised By**

**Name:** Mr Simon Parks

**Title:** Senior Biosafety Scientist