



Certificate of Containment Testing

Containment Testing of Rotor R22A4 in a Floor Standing Centrifuge

Report No. 40/13 A

Report Prepared For: Eppendorf Himac Technologies Co., Ltd.

Issue Date: 14 May 2021

Test Summary

Rotor R22A4 was containment tested in the CR22 GIII floor standing centrifuge, using Annex AA of IEC 61010-2-020:2006 (2nd Ed.). The sealed rotor was shown to contain a spill within the rotor.

Report Written By

Name: Ms Helen Hookway

Title: Biosafety Scientist

Report Authorised By

Name: Mrs Sara Speight

Title: Senior Biosafety Scientist