



Public Health  
England

Public Health England  
Microbiology Services  
Porton Down  
Salisbury  
Wiltshire  
SP4 OJG

# Certificate of Containment Testing

## Containment Testing of Rotor R5S4 in a Floor Standing Centrifuge

Report No. 40/13 D

**Report Prepared For:** Eppendorf Himac Technologies Co., Ltd.

**Issue Date:** 14 May 2021

### Test Summary

Rotor R5S4 was containment tested in the CR22GIII floor standing centrifuge, using Annex AA of IEC 61010-2-020:2006 (2<sup>nd</sup> Ed.). The sealed rotor was shown to contain a spill within the rotor.

**Report Written By**

**Name:** Ms Helen Hookway

**Title:** Biosafety Scientist

**Report Authorised By**

**Name:** Mrs Sara Speight

**Title:** Senior Biosafety Scientist