



Public Health  
England

Public Health England  
Microbiology Services  
Porton Down  
Salisbury  
Wiltshire  
SP4 OJG

# Certificate of Containment Testing

## Containment Testing of Rotor R9A in a Floor Standing Centrifuge

Report No. 40/13 C

**Report Prepared For:** Eppendorf Himac Technologies Co., Ltd.

**Issue Date:** 14 May 2021

### Test Summary

Rotor R9A was containment tested in the CR22GIII floor standing centrifuge, using Annex AA of IEC 61010-2-020:2006 (2<sup>nd</sup> Ed.). The sealed rotor was shown to contain a spill within the rotor.

#### Report Written By

**Name:** Ms Helen Hookway

**Title:** Biosafety Scientist

#### Report Authorised By

**Name:** Mrs Sara Speight

**Title:** Senior Biosafety Scientist

Please be aware that the use of the Royal Coat of Arms is highly restricted and cannot be copied. Please do not put the PHE logo on your website or use our name to endorse your products. Any reference to PHE needs to be approved by us before it can be used.