

# CERTIFICATE OF COMPLIANCE

**Certificate Number** 2023-02-24-E215059  
**Report Reference** E215059-D1027-1/A0/C0-ULCB  
**Date** 2023-02-24

**Issued to:** EPPENDORF SE  
**Applicant Company:** BARKHAUSENWEG 1  
22339 HAMBURG GERMANY

**Listed Company:** Same as Applicant

**This is to certify that  
representative samples of** Laboratory centrifuge  
5427R (5429)

Have been investigated by UL in accordance with the  
Standard(s) indicated on this Certificate.

**Standard(s) for Safety:** UL 61010-1, 3rd Edition, May 11, 2012, Revised July 19, 2019,  
CAN/CSA-C22.2 No. 61010-1(2012-05), 3rd Edition, with  
revisions through 2018-11

**Additional Standards:** IEC 61010-2-011:2019, UL 61010-2-011:2021, CSA-C22.2 No.  
61010-2-011:19;  
IEC 61010-2-020:2016, CAN/CSA-C22.2 No. 61010-2-020:17;

Only the 220-240 V device was evaluated for the IEC 61010-2-  
101: 2018 (Third Edition) (CB report)

**Additional Information:** See the UL Online Certifications Directory at  
[www.ul.com/database](http://www.ul.com/database) for additional information.

This Certificate of Compliance does not provide authorization to apply the UL Mark. Only the UL Follow-Up Services Procedure provides authorization to apply the UL Mark.

Only those products bearing the UL Mark should be considered as being UL Certified and covered under UL's Follow-Up Services.

Look for the UL Certification Mark on the product.

*B. Mahlen*

Bruce Mahrenholz, Director North American Certification Program  
UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please  
contact a local UL Customer Service Representative at <http://ul.com/aboutul/locations/>

